



Report SEM / EDX: RP_5698_TA-Dent_617529_T4-Switch_3,3x12_asm

Client: TA-Dent GmbH
Dr. Tilmann Beck
Moltkestrasse 78
77654 Offenburg - Germany

Date of Order: 05 August 2016

Concern: SEM analysis - implant / abutment connection

Speciemen designation: T4-Switch 3,3x12,0 implant with TA-Dent abutment D4 Gh2 A0
implant LOT: 617529 implant REF: K154-01-33012
abutment LOT: 38721 abutment REF: K154-70-102040

Order number: 5698

Date of examination: 25 August 2016

Examination specification: AA-Diagnostics_02
Display of specimen surfaces via secondary electrons (SE) - SEM.

The following should be taken into consideration when examining the spectra and interpreting the results (only for EDX analysis):

Overlapping of lines:

Lines of different chemical elements, which are located in the spectra close together, overlap on lower deviation of EDX detectors dissolution of energy. Partially line overlapping become dissolved by expansion of the spectra. Element selection takes place on the basis of experience for the probably arising element.

Detection limit:

Due to the actuating voltage and other system parameters, like for example acceleration voltage and beam current, the detection of element concentrations lower than 0,1 masspercent is impossible.

Depth of information:

The information depth which causes the largest part of the analysis signal, depends on the penetration depth of the electron beam and on the distance which the x-ray beam can go through the sample. It reaches from one to some micrometers.

Quantification:

To the quantification of the speciemen - due to the rough surface - the fundamental-parameter-based quantitative P/U-ZAF method is used, which is based on the evaluation of the characteristic x-ray and the brake radiation underground. The local peak-to-underground conditions (P/U) are received into the modified ZAF matrix correction.

Examination results: See following pages.

25 August 2016


Christoph Stoll
- General Manager -

Report SEM / EDX:

RP_5698_TA-Dent_617529_T4-Switch_3,3x12_asm

State of specimen:

unsterile assembly (see figure 1-4)

Examination procedure / equipment:

The examination is done with a Zeiss EVO MA15 REM (Software SmartSEM V05.04.03.00 16-Apr.-10) under high vacuum conditions. The high voltage of the SEM is set to 15keV with a cathode current of 60pA for high resolution image or 1000pA for EDX scanning. **(AA-Diagnostics_02)**
The element analysis of the examination surface is done with a BRUKER X-Flash 5010 EDX detector, equipped with a light element window type slew AP3.3. (Software Esprit 1.9 - Quantax 200). The slew AP3.3. allows the detection of elements with an ordinal number ≥ 5 (Boron).



Figure no. 1: components of the assembly



Figure no. 2: embeded assembly




Figure no. 3: embeded assembly in SEM table




Figure no. 4: SEM survey - embeded assembly 30x

Report SEM / EDX:

RP_5698_TA-Dent_617529_T4-Switch_3,3x12_asm




Figure no. 5: SE image - implant / abutment / fixation screw; magnification: 50x




Figure no. 6: SE image - implant / abutment / fixation screw; magnification: 100x



Figure no. 7: SE image - detail implant / abutment gap - see figure no. 6; magnification: 500x

The results of the report only advert to the examined specimens. The use of this report in extracts is not permitted!

Report SEM / EDX:

RP_5698_TA-Dent_617529_T4-Switch_3,3x12_asm




Figure no. 8: SE image - abutment / fixation screw; magnification: 50x



Figure no. 9: SE image - detail - head of the fixation screw - see figure no. 8; magnification: 1000x




Figure no. 10: SE image - survey of the thread (implant and fixation screw); magnification: 100x

The results of the report only advert to the examined specimens. The use of this report in extracts is not permitted!